

## SUPPLEMENTAL / BID BULLETIN NO. 2

Title : **Supply and Delivery of Weighing Scales, Measuring Tapes and Wooden Height Boards for DSWD BangUN Project**

ITB No. : **GOP/18-DSWD-019**

Date : **10 August 2018**

---

This Supplemental/ Bid Bulletin is issued to all prospective bidders announcing changes in the Bidding Documents, to wit:

### I. Section VI. Schedule of Requirements

Particulars	From	To
Delivery Period	Complete delivery of goods must be made within thirty (30) calendar days upon receipt of Notice to Proceed (NTP)	<b>Complete delivery of goods must be made within fifty (50) calendar days upon receipt of Notice to Proceed (NTP)</b>

### II. Section VII. Technical Specification

Particulars	From	To
Lot No. 1 Electronic/Digital Weighing Scale	4. Dimension (WxHxD) : $\leq 316$ mm x $\leq 37$ mm x $326$ mm  10. Every unit should be DOST Calibrated.	<b>4. Dimension (WxHxD) : <math>\leq 316</math> mm x <math>\leq 37</math> mm x <math>\leq 326</math> mm</b>  <b>10. Random calibration by DOST</b>
Lot No. 2 Mechanical Measuring Tape	6. Dimension (WxHxD) : $\leq 125$ mm x $\leq 125$ mm x $173$ mm	<b>6. Dimension (WxHxD) : <math>\leq 125</math> mm x <math>\leq 125</math> mm x <math>\leq 173</math> mm</b>
Lot No. 3 Hanging-type Weighing Scale	10. Every unit should be DOST Calibrated.	<b>10. Random calibration by DOST</b>
Lot No. 4 Wooden Height Board	3. Location of Ruler : any place in the board  5. Every unit should be DOST Calibrated.	<b>3. Location of Ruler : middle of the height board (see attached picture)</b>  <b>5. Random calibration by DOST</b>

**Question:**

1. Will the DSWD consider “any Philippine Accreditation Board-accredited laboratories that calibrates human weight” as additional calibrating agencies beside DOST for calibrating the items?

**Response:**

**Only DOST-calibrated items will be accepted.**

**Question:**

2. What is the accuracy level or marginal error of the weighing scale for calibration?

**Response:**

**The accuracy level or marginal error that is acceptable is 0.1 to 0.3.**

**Question:**

3. Will the DSWD consider the submission of sample products during the opening of bids?

**Response:**

**Sample products will be submitted by the Lowest Calculated Bid (LCB) during post-qualification as stipulated in Section 29.2 of the Bidding Documents.**

### **III. Deadline for the Submission and Receipt of Bids**

This is to reiterate that the deadline for the submission and receipt of bids is **17 August 2018** at **01:00 p.m.** Bids must be delivered to the BAC Secretariat Office, Ground Floor, DSWD Central Office. **Late Bids will not be accepted.**

The Bid Opening will immediately follow after the deadline, to be held at the Legal Service Conference Room, 4<sup>th</sup> Floor, DSWD Central Office, IBP Road, Constitution Hills, Quezon City.

### **IV. Issuance of Bidding Documents**

The issuance of bidding documents is extended until **17 August 2018** at **12:00 p.m.**

This Bid Bulletin shall form part of the bidding documents.

Please be guided accordingly.

**LUZVIMINDA C. ILAGAN**  
Undersecretary and  
Bids and Awards Committee Chairperson

By:

*(Original Signed)*  
**MARIE ANGELA S. GOPALAN**  
Director IV and  
Bids and Awards Committee Vice-Chairperson

**Please accomplish this portion and send it to facsimile no. (02) 951-7116 or email at bacsec@dswd.gov.ph**

---

Received from DSWD, Supplemental/Bid Bulletin No. 2 for the **Supply and Delivery of Weighing Scales, Measuring Tapes and Wooden Height Boards for DSWD BangUN Project** (ITB No. GOP/18-DSWD-019).

Received by : \_\_\_\_\_ Date : \_\_\_\_\_  
(Signature over Printed Name)

Designation : \_\_\_\_\_

Company : \_\_\_\_\_

**HEIGHT BOARD WITH STEEL RULER**



**FRONT MID RANGE**



**BACK AND SIDE VIEW**